



For Research Use Only, Not for use in diagnostic procedures

ELISA Kit for Measuring Mouse FABP5/E-FABP/mal1

CircuLex Mouse FABP5/E-FABP/mal1 ELISA Kit

Cat# CY-8056

Intended Use..... 1
 Storage..... 1
 Introduction..... 2
 Principle of the Assay..... 3
 Materials Provided..... 4
 Materials Required but not Provided..... 4
 Precautions and Recommendations..... 5
 Sample Collection and Storage..... 6
 Detailed Protocol..... 7-9
 Calculations..... 9
 Measurement Range..... 9
 Troubleshooting..... 9
 Reagent Stability..... 9-10
 Assay Characteristics..... 10-12
 Example of Test Results..... 12
 References..... 13

Intended Use

The MBL Research Product **CircuLex Mouse FABP5/E-FABP/mal1 ELISA Kit** is used for the quantitative measurement of mouse FABP5/E-FABP/mal1 in serum.

Individual users should determine appropriate conditions when using other types of samples.

This assay kit is for research use only and not for use in diagnostic or therapeutic procedures.

Storage

- Upon receipt store all components at 4°C.
- Don't expose reagents to excessive light.

Introduction


Fatty acid-binding proteins (FABPs) are a class of cytoplasmic proteins that bind long chain fatty acids. FABPs are small intracellular proteins (~13-14 kDa) with a high degree of tissue specificity. They are abundantly present in various cell types and play an important role in the intracellular utilization of fatty acids, transport and metabolism. There are at least nine distinct types of FABP, each showing a specific pattern of tissue expression.

Human epidermal fatty acid binding protein (E-FABP, FABP5 and in mice called also mal1) belongs to the intracellular FABP family and is expressed in differentiated adipocytes, macrophages, skin, brain, mammary glands and so on. Mouse FABP5/E-FABP/mal1 has a high degree of homology to aP2 (mouse FABP4) and is also an important player in obesity-related disorders. FABP5/E-FABP/mal1 knockout mice exhibited enhanced insulin-stimulated glucose uptake and increased systemic insulin sensitivity, while transgenic overexpression of FABP5/E-FABP/mal1 aggravated insulin resistance and hyperglycaemia. Conversely, a marked compensatory up-regulation of adipocyte FABP5/E-FABP/mal1 expression was observed in FABP4/ aP2-deficient mice.

Principle of the Assay

The MBL Research Product **CircuLex Mouse FABP5/E-FABP/mal1 ELISA Kit** employs the quantitative sandwich enzyme immunoassay technique. An antibody specific for mouse FABP5/E-FABP/mal1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and the immobilized antibody binds any mouse FABP5/E-FABP/mal1 present. After washing away any unbound substances, a biotinylated antibody specific for mouse FABP5/E-FABP/mal1 is added to the wells, followed by binding with horseradish peroxidase (HRP) conjugated Streptavidin, which then catalyzes the conversion of the chromogenic substrate tetra-methylbenzidine (TMB) from a colorless solution to a blue solution (or yellow after the addition of stopping reagent). The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured at 450 nm. The absorbance is proportional to the concentration of mouse FABP5/E-FABP/mal1. A standard curve is constructed by plotting absorbance values versus mouse FABP5/E-FABP/mal1 concentrations of calibrators, and concentrations of unknown samples are determined using this standard curve.

Summary of Procedure



Materials Provided

All samples and standards should be assayed in duplicate. The following components are supplied and are sufficient for the one 96-well microplate kit.

Microplate: One microplate supplied ready to use, with 96 wells (12 strips of 8-wells) in a foil, zip-lock bag with a desiccant pack. Wells are coated with anti-mouse FABP5/E-FABP/mal1 antibody as a capture antibody.

10X Wash Buffer: One bottle containing 100 mL of 10X buffer containing Tween®-20

Dilution Buffer: One bottle containing 50 mL of 1X buffer; use for sample dilution. Ready to use.

Mouse FABP5/E-FABP/mal1 Standard: One vials containing X* ng of lyophilized recombinant mouse FABP5/E-FABP/mal1

***The amount is changed depending on lot. See the real "User's Manual" included in the kit box.**

Biotinylated Detection Antibody: One bottle containing 12 mL of biotinylated anti-mouse FABP5/E-FABP/mal1 antibody. Ready to use.

HRP conjugated Streptavidin: One bottle containing 12 mL of HRP (horseradish peroxidase) conjugated streptavidin. Ready to use.

Substrate Reagent: One bottle containing 20 mL of the chromogenic substrate, tetra-methylbenzidine (TMB). Ready to use.

Stop Solution: One bottle containing 20 mL of 1 N H₂SO₄. Ready to use.

Materials Required but not Provided

- **Pipettors:** 2-20 µL, 20-200 µL and 200-1,000 µL precision pipettors with disposable tips.
- **Precision repeating pipettor**
- **Orbital microplate shaker**
- **Microcentrifuge and tubes** for sample preparation.
- **Vortex mixer**
- **(Optional) Microplate washer:** Manual washing is possible but not preferable.
- **Plate reader** capable of measuring absorbance in 96-well plates at dual wavelengths of 450 nm/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. The plate can also be read at a single wavelength of 450 nm, which will give a somewhat higher reading.
- **(Optional) Software package facilitating data generation and analysis**
- **500 or 1,000 mL graduated cylinder.**
- **Reagent reservoirs**
- **Deionized water of the highest quality**
- **Disposable paper towels**

Precautions and Recommendations

- **Although we suggest to conduct experiments as outlined below, the optimal experimental conditions will vary depending on the parameters being investigated, and must be determined by the individual user.**
- Allow all the components to come to room temperature before use.
- All microplate strips that are not immediately required should be returned to the zip-lock pouch, which must be carefully resealed to avoid moisture absorption.
- Do not use kit components beyond the indicated kit expiration date.
- Use only the microtiter wells provided with the kit.
- Rinse all detergent residue from glassware.
- Use deionized water of the highest quality.
- Do not mix reagents from different kits.
- The buffers and reagents in this kit may contain preservatives or other chemicals. Care should be taken to avoid direct contact with these reagents.
- Do not mouth pipette or ingest any of the reagents.
- Do not smoke, eat, or drink when performing the assay or in areas where samples or reagents are handled.
- Dispose of tetra-methylbenzidine (TMB) containing solutions in compliance with local regulations.
- Avoid contact with Substrate Solution which contains hydrogen peroxide.
- Avoid contact with Stop Solution which contains Sulfuric Acid.
- Wear gloves and eye protection when handling immunodiagnostic materials and samples of human origin, and these reagents. In case of contact with the Stop Solution and the Substrate Solution, wash skin thoroughly with water and seek medical attention, when necessary.
- **Biological samples may be contaminated with infectious agents. Do not ingest, expose to open wounds or breathe aerosols. Wear protective gloves and dispose of biological samples properly.**
- **CAUTION: Sulfuric Acid is a strong acid. Wear disposable gloves and eye protection when handling Stop Solution.**

Sample Collection and Storage

Serum: Use a serum separator tube and allow samples to clot for 60 ± 30 minutes. Centrifuge the samples at 4°C for 10 minutes at 1,000 x g. Remove serum and assay immediately or store samples on ice for up to 6 hours before assaying. Aliquots of serum may also be stored at below -70°C for extended periods of time. Avoid repeated freeze-thaw cycles.

Other biological samples: MBL has not tested.

(*e.g.* Remove any particulates by centrifugation and assay immediately or aliquot and store samples at below -70°C. Avoid repeated freeze-thaw cycles. Individual users should determine appropriate conditions when using other types of samples.)

Detailed Protocol

The MBL Research Product **CircuLex Mouse FABP5/E-FABP/mal1 ELISA Kit** is provided with removable strips of wells so the assay can be carried out on separate occasions using only the number of strips required for the particular determination. Since experimental conditions may vary, an aliquot of the mouse FABP5/E-FABP/mal1 Standard within the kit, should be included in each assay as a calibrator. Disposable pipette tips and reagent troughs should be used for all liquid transfers to avoid cross-contamination of reagents or samples.

Preparation of Working Solutions

All reagents need to be brought to room temperature prior to the assay. Assay reagents are supplied ready-to-use, with the exception of **10X Wash Buffer** and **Mouse FABP5/E-FABP/mal1 Standard**.

1. Prepare a working solution of Wash Buffer by adding 100 mL of **10X Wash Buffer** to 900 mL of deionized (distilled) water (**ddH₂O**). Mix well. Store at 4°C for two weeks or -20°C for long-term storage.
2. Reconstitute **Mouse FABP5/E-FABP/mal1 Standard** with **X* mL** of **ddH₂O** by gently mixing. After reconstitution, immediately dispense it in small aliquots (e.g. 80 µL) to plastic micro-centrifuge tubes and store below -70°C to avoid non-specific adsorption to glass surface and multiple freeze-thaw cycles. The concentration of the mouse FABP5/E-FABP/mal1 in vial should be **16 ng/mL**, which is referred to as the **Master Standard** of mouse FABP5/E-FABP/mal1.

***The amount is changed depending on lot. See the real "User's Manual" included in the kit box.**

Prepare Standard Solutions as follows:

Use the **Master Standard** to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 4,000 pg/mL standard (Std.1) serves as the highest standard. The **Dilution Buffer** serves as the zero standard (Blank).

	Volume of Standard	Dilution Buffer	Concentration
Std.1	150 µL of Master Standard	450 µL	4,000 pg/mL
Std.2	300 µL of Std. 1 (4,000 pg/mL)	300 µL	2,000 pg/mL
Std.3	300 µL of Std. 2 (2,000 pg/mL)	300 µL	1,000 pg/mL
Std.4	300 µL of Std. 3 (1,000 pg/mL)	300 µL	500 pg/mL
Std.5	300 µL of Std. 4 (500 pg/mL)	300 µL	250 pg/mL
Std.6	300 µL of Std. 5 (250 pg/mL)	300 µL	125 pg/mL
Std.7	300 µL of Std. 6 (125 pg/mL)	300 µL	62.5 pg/mL
Blank	-	300 µL	0 pg/mL

Note: Do not use a Repeating pipette. Change tips for every dilution. Wet tip with Dilution Buffer before dispensing. Discard any unused Standard Solutions after use.

Sample Preparation

Dilute samples with **Dilution Buffer**.

- Serum samples may require a 200-600 fold dilution with Dilution Buffer.

Assay Procedure

1. Remove the appropriate number of microtiter wells from the foil pouch and place them into the well holder. Return any unused wells to the foil pouch, refold, seal with tape and store at 4°C.
2. Dilute samples with **Dilution Buffer**. (See “Sample Preparation” above.)
3. Pipette **100 µL** of **Standard Solutions (Std1-Std7, Blank)** and **diluted samples** in duplicates, into the appropriate wells.
4. Incubate the plate **at room temperature (ca.25°C) for 60 minutes** shaking at ca. 300 rpm on an orbital microplate shaker.
5. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
6. Add **100 µL** of **Biotinylated Detection Antibody** into each well.
7. Incubate the plate **at room temperature (ca.25°C) for 60 minutes**, shaking at ca. 300 rpm on an orbital microplate shaker.
8. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
9. Add **100 µL** of **HRP conjugated Streptavidin** into each well.
10. Incubate the plate **at room temperature (ca.25°C) for 20 minutes**, shaking at ca. 300 rpm on an orbital microplate shaker.
11. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
12. Add **100 µL** of **Substrate Reagent**. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with e.g. aluminum foil is recommended. Return Substrate A to 4°C immediately after the necessary volume is removed
13. Incubate the plate **at room temperature (ca. 25°C) for 10-20 minutes**, shaking at ca. 300 rpm on an orbital microplate shaker. The incubation time may be extended up to 30 minutes if the reaction temperature is below than 20°C.
14. Add **100 µL** of **Stop Solution** to each well in the same order as the previously added Substrate Reagent.
15. Measure absorbance in each well using a spectrophotometric microplate reader at dual wavelengths of 450/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. Read the microplate at 450 nm if only a single wavelength can be used. Wells must be read within 30 minutes of adding the Stop Solution.

Note-1: Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

Note-2: Reliable standard curves are obtained when either O.D. values do not exceed 0.2 units for the blank (zero concentration), or 2.5 units for the highest standard concentration.

Note-3: If the microplate reader is not capable of reading absorbance greater than the absorbance of the highest standard, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine mouse FABP5/E-FABP/mal1 concentration of off-scale samples. The readings at 405 nm should not replace the on-scale readings at 450 nm.

Calculations

Average the duplicate readings for each standard, control and sample, and subtract the optical density of the average zero standard. Plot the optical density versus the concentration of standards and draw the best curve. Most microtiter plate readers perform automatic calculations of analyte concentration. The standard curve fits best to a sigmoidal four-parameter logistic equation. The results of unknown samples can be calculated with any computer program having a four-parameter logistic function.

A standard curve is also to be constructed by plotting the absorbance (Y) versus log of the known concentration (X) of standards, using a cubic function. Alternatively, the logit log function can be used to linearize the standard curve (i.e. logit of optical density (Y) is plotted versus log of the known concentration (X) of standards). To determine the concentration of each sample, first find the optical density on the y-axis and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the x-axis and read the corresponding concentration.

If the samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Measurement Range

The measurement range is 62.5 pg/mL to 4,000 pg/mL. Any sample reading higher than the highest standard should be diluted with Dilution Buffer in higher dilution and re-assayed. Dilution factors need to be taken into consideration in calculating the mouse FABP5/E-FABP/mal1 concentration.

Troubleshooting

1. All samples and controls should be assayed in duplicate, using the protocol described in the **Detailed Protocol**. Incubation times or temperatures significantly different from those specified may give erroneous results.
2. Poor duplicates, accompanied by elevated values for wells containing no sample, indicate insufficient washing. If all instructions in the **Detailed Protocol** were followed accurately, such results indicate a need for washer maintenance.
3. Overall low signal may indicate that desiccation of the plate has occurred between the final wash and addition of Substrate Reagent. Do not allow the plate to dry out. Add Substrate Reagent immediately after wash.

Reagent Stability

All of the reagents included in the MBL Research Product **CircuLex Mouse FABP5/E-FABP /mal1 ELISA Kit** have been tested for stability. Reagents should not be used beyond the stated expiration date. Upon receipt, kit reagents should be stored at 4°C, except the reconstituted mouse FABP5/E-FABP/mal1

Standard must be stored at below -70°C. Coated assay plates should be stored in the original foil bag sealed by the zip lock and containing a desiccant pack.


Assay Characteristics

1. Sensitivity

The limit of detection (defined as such a concentration of mouse FABP5/E-FABP/mal1 giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: A blank + 3SD blank) is better than 6.91 pg/ml of sample.

* Dilution Buffer was pipetted into blank wells.

Typical standard curve



2. Precision

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested sixteen times on one plate to assess intra-assay precision.

- Intra-assay (Within-Run, n=16) CV=4.8-5.0 %

Mouse FABP 5 conc. (ng/ml)			
	Serum 1	Serum 2	Serum 3
1	395.80	150.97	55.60
2	396.24	158.16	54.43
3	398.58	160.12	49.51
4	418.47	166.12	53.13
5	417.68	171.32	58.40
6	422.58	165.61	55.82
7	396.36	168.48	55.34
8	381.89	167.26	60.37
9	366.38	157.46	57.35
10	405.98	154.18	55.18
11	427.35	146.07	50.29
12	424.08	151.60	56.41
13	422.38	150.98	55.92
14	428.64	159.15	55.62
15	391.39	162.85	56.63
16	371.41	172.33	54.72
MAX.	428.64	172.33	60.37
MIN.	366.38	146.07	49.51
MEAN	404.08	160.17	55.29
S.D.	19.97	7.97	2.68
C.V.	4.9%	5.0%	4.8%

Inter-assay Precision (Precision between assays)


Three samples of known concentration were tested in four separate assays to assess inter-assay precision.

- Inter-assay (Run-to-Run, n=4) CV=2.6-9.9 %

Mouse FABP 5 conc. (ng/ml)			
	Serum 1	Serum 2	Serum 3
1	507.2	170.6	52.2
2	508.9	208.3	53.9
3	485.2	213.4	43.6
4	515.7	191.1	54.0
MAX.	515.7	213.4	54.0
MIN.	485.2	170.6	43.6
MEAN	504.3	195.9	50.9
S.D.	13.25	19.33	4.94
C.V.	2.63%	9.87%	9.71%

3. Linearity


Three samples* were diluted with Dilution Buffer and assayed after dilution. The neat sample was set to 1.



* Sample : Serum

Example of Test Results

Fig. FABP5/E-FABP concentration in mouse sera (N=22).



References

1. Furuhashi M and Hotamisligil GS. Fatty acid-binding proteins: role in metabolic diseases and potential as drug targets. *Nat Rev Drug Discov.* 7, 489 (2008)
2. Chmurzynska A. The multigene family of fatty acid-binding proteins (FABPs): function, structure and polymorphism. *J Appl Genet.* 47, 39-48 (2006)
3. Storch J and Corsico B. The emerging functions and mechanisms of mammalian fatty acid-binding proteins. *Annu Rev Nutr.* 28,73-95 (2008)
4. Krieg P *et al.* Tumor-specific overexpression of a novel keratinocyte lipid-binding protein. Identification and characterization of a cloned sequence activated during multistage carcinogenesis in mouse skin. *J Biol Chem.* 268,17362-9 (1993)
5. Siegenthaler G *et al.* Purification and characterization of the human epidermal fatty acid-binding protein: localization during epidermal cell differentiation in vivo and in vitro. *Biochem J.* 302, 363-71 (1994)
6. Maeda K *et al.* Role of the fatty acid binding protein mal1 in obesity and insulin resistance. *Diabetes.* 52, 300-7 (2003)
7. Hotamisligil GS *et al.* Uncoupling of obesity from insulin resistance through a targeted mutation in aP2, the adipocyte fatty acid binding protein. *Science.* 274, 1377-9 (1996)
8. Scheja L *et al.* Altered insulin secretion associated with reduced lipolytic efficiency in aP2^{-/-} mice. *Diabetes.* 48,1987-94 (1999)
9. Shaughnessy S *et al.* Adipocyte metabolism in adipocyte fatty acid binding protein knockout mice (aP2^{-/-}) after short-term high-fat feeding: functional compensation by the keratinocyte fatty acid binding protein. *Diabetes.* 49, 904-11 (2000)

For more information, please visit our website at <https://ruo.mbl.co.jp/>.

MANUFACTURED BY



URL: <https://ruo.mbl.co.jp>
E-mail: support@mbl.co.jp

CycLex/CircuLex products are supplied for research use only. CycLex/CircuLex products and components thereof may not be resold, modified for resale, or used to manufacture commercial products without prior written approval from MBL. To inquire about licensing for such commercial use, please contact us via email.