

ELISA Kit for Measuring Human soluble LOX-1

CircuLex Human soluble LOX-1/OLR1 ELISA Kit

Cat# CY-8081

Intended Use.....	1
Storage.....	1
Introduction.....	2
Principle of the Assay.....	2-3
Materials Provided.....	3
Materials Required but not Provided.....	4
Precautions and Recommendations.....	5
Sample Collection and Storage.....	6
Detailed Protocol.....	7-8
Calculations.....	9
Measurement Range.....	9
Troubleshooting.....	9
Reagent Stability.....	9
Assay Characteristics.....	10-12
Example of Test Results.....	13
References.....	14

Intended Use

The MBL Research Product **CircuLex Human soluble LOX-1/OLR1 ELISA Kit** is used for the quantitative measurement of human soluble LOX-1 in cell culture supernatant.

Individual users should determine appropriate conditions when using other types of samples.

This assay kit is for research use only and not for use in diagnostic or therapeutic procedures.

Storage

- Upon receipt store all components at 4°C.
- Don't expose reagents to excessive light.

Introduction


Lectin-like oxidized LDL receptor-1 (LOX-1), a 52-kD type II transmembrane receptor for oxidized low-density lipoproteins (ox-LDL) belonging to the C-type lectin family is present primarily on endothelial cells (1). Accumulating evidences indicate that oxLDL uptake through this receptor induces endothelial dysfunction. oxLDL binding to endothelial LOX-1 generates superoxide anions, decreases nitric oxide production, and activates NF-kB (2, 3). Endothelial LOX-1 expression is induced by various pro-inflammatory cytokines, such as TNF-alpha (4) and TGF-beta (5) as well as by pro-atherogenic factors, such as oxLDL and advanced glycation end products in vitro (6). This receptor is expressed in the aortas of hypertensive, diabetic, and hyperlipidemic animals (6-8) and is upregulated in early human atherosclerotic lesions (9). These results suggest that LOX-1 may be expressed locally and play important roles in atherogenesis by internalizing and degrading oxLDL and in inflammatory responses in vivo. LOX-1 can be cleaved from the cell surface and released as soluble LOX-1 (sLOX-1), and elevated sLOX-1 levels may be indicative of atherosclerotic plaque instability (10).

Principle of the Assay

The MBL Research Product **CircuLex Human soluble LOX-1/OLR1 ELISA Kit** employs the quantitative sandwich enzyme immunoassay technique. An antibody specific for human soluble LOX-1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and the immobilized antibody binds any human soluble LOX-1 present. After washing away any unbound substances, an HRP conjugated antibody specific for human soluble LOX-1 is added to the wells. Following a wash to remove any unbound HRP conjugated antibody, the remaining conjugate is allowed to react with the substrate H₂O₂-tetramethylbenzidine. The reaction is stopped by addition of acidic solution and absorbance of the resulting yellow product is measured at 450 nm. The absorbance is proportional to the concentration of human soluble LOX-1. A standard curve is constructed by plotting absorbance values versus human soluble LOX-1 concentrations of calibrators, and concentrations of unknown samples are determined using this standard curve.

The MBL Research Product **CircuLex Human soluble LOX-1/OLR1 ELISA Kit** is designed to measure the concentration of human soluble LOX-1 in conditioned medium or human serum/plasma.

Summary of Procedure



Materials Provided

All samples and standards should be assayed in duplicate. The following components are supplied and are sufficient for the one 96-well microplate kit.

Microplate: One microplate supplied ready to use, with 96 wells (12 strips of 8-wells) in a foil, zip-lock bag with a desiccant pack. Wells are coated with anti-human soluble LOX-1 antibody as a capture antibody.

10X Wash Buffer: One bottle containing 100 mL of 10X buffer containing Tween® -20

Dilution Buffer: One bottle containing 50 mL of 1X buffer; use for reconstitution of human soluble LOX-1 Standard and sample dilution. Ready to use.

Human soluble LOX-1 Standard: One vial containing X* ng of lyophilized recombinant human soluble LOX-1

***The amount is changed depending on lot. See the real "User's Manual" included in the kit box.**

HRP conjugated Detection Antibody: One bottle containing 12 mL of HRP (horseradish peroxidase) conjugated anti-human soluble LOX-1 antibody. Ready to use.

Substrate Reagent: One bottle containing 20 mL of the chromogenic substrate, tetra-methylbenzidine (TMB). Ready to use.

Stop Solution: One bottle containing 20 mL of 1 N H₂SO₄. Ready to use.

Materials Required but not Provided

- **Pipettors:** 2-20 μ L, 20-200 μ L and 200-1,000 μ L precision pipettors with disposable tips.
- **Precision repeating pipettor**
- **Orbital microplate shaker**
- **Microcentrifuge and tubes** for sample preparation.
- **Vortex mixer**
- **(Optional) Microplate washer:** Manual washing is possible but not preferable.
- **Plate reader** capable of measuring absorbance in 96-well plates at dual wavelengths of 450 nm/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. The plate can also be read at a single wavelength of 450 nm, which will give a somewhat higher reading.
- **(Optional) Software package facilitating data generation and analysis**
- **500 or 1,000 mL graduated cylinder.**
- **Reagent reservoirs**
- **Deionized water of the highest quality**
- **Disposable paper towels**

Precautions and Recommendations

- **Although we suggest to conduct experiments as outlined below, the optimal experimental conditions will vary depending on the parameters being investigated, and must be determined by the individual user.**
- Allow all the components to come to room temperature before use.
- All microplate strips that are not immediately required should be returned to the zip-lock pouch, which must be carefully resealed to avoid moisture absorption.
- Do not use kit components beyond the indicated kit expiration date.
- Use only the microtiter wells provided with the kit.
- Rinse all detergent residue from glassware.
- Use deionized water of the highest quality.
- Do not mix reagents from different kits.
- The buffers and reagents in this kit may contain preservatives or other chemicals. Care should be taken to avoid direct contact with these reagents.
- Do not mouth pipette or ingest any of the reagents.
- Do not smoke, eat, or drink when performing the assay or in areas where samples or reagents are handled.
- Dispose of tetra-methylbenzidine (TMB) containing solutions in compliance with local regulations.
- Avoid contact with the acidic Stop Solution and Substrate Solution, which contains hydrogen peroxide.
- Wear gloves and eye protection when handling immunodiagnostic materials and samples of human origin, and these reagents. In case of contact with the Stop Solution and the Substrate Solution, wash skin thoroughly with water and seek medical attention, when necessary.
- **Biological samples may be contaminated with infectious agents. Do not ingest, expose to open wounds or breathe aerosols. Wear protective gloves and dispose of biological samples properly.**
- **CAUTION: Sulfuric Acid is a strong acid. Wear disposable gloves and eye protection when handling Stop Solution.**

Sample Collection and Storage

Cell culture supernatant: Remove any particulates by centrifugation and assay immediately or aliquot and store samples below -70°C . Avoid repeated freeze-thaw cycles.

Other biological samples: MBL has not tested.

(e.g. Remove any particulates by centrifugation and assay immediately or aliquot and store samples at below -70°C . Avoid repeated freeze-thaw cycles. Individual users should determine appropriate conditions when using other types of samples.)

Detailed Protocol

The MBL Research Product **CircuLex Human soluble LOX-1/OLR1 ELISA Kit** is provided with removable strips of wells so the assay can be carried out on separate occasions using only the number of strips required for the particular determination. Since experimental conditions may vary, an aliquot of the human soluble LOX-1 Standard within the kit, should be included in each assay as a calibrator. Disposable pipette tips and reagent troughs should be used for all liquid transfers to avoid cross-contamination of reagents or samples.

Preparation of Working Solutions

All reagents need to be brought to room temperature prior to the assay. Assay reagents are supplied ready-to-use, with the exception of **10X Wash Buffer** and **Human soluble LOX-1 Standard**.

1. Prepare a working solution of Wash Buffer by adding 100 mL of the **10X Wash Buffer** to 900 mL of deionized (distilled) water (ddH₂O). Mix well. Store at 4°C for two weeks or -20°C for long-term storage.
2. Reconstitute **Human soluble LOX-1 Standard** with **X* mL** of **Dilution Buffer** by gently mixing. After reconstitution, immediately dispense it in small aliquots (e.g. 100 µL) to plastic micro-centrifuge tubes and store below -70°C to avoid non-specific adsorption to glass surface and multiple freeze-thaw cycles. The concentration of the reconstituted Human soluble LOX-1 Standard should be **24 ng/mL**, which is referred to as the **Master Standard** of human soluble LOX-1.

***The amount is changed depending on lot. See the real "User's Manual" included in the kit box.**

Prepare Standard Solutions as follows:

Use the **Master Standard** to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 2,400 pg/mL standard (Std.1) serves as the highest standard. The **Dilution Buffer** serves as the zero standard (Blank).

	Volume of Standard	Dilution Buffer	Concentration
Std.1	60 µL of Master Standard	540 µL	2,400 pg/mL
Std.2	300 µL of Std. 1 (2,400 pg/ml)	300 µL	1,200 pg/mL
Std.3	300 µL of Std. 2 (1,200 pg /ml)	300 µL	600 pg/mL
Std.4	300 µL of Std. 3 (600 pg /ml)	300 µL	300 pg/mL
Std.5	300 µL of Std. 4 (300 pg /ml)	300 µL	150 pg/mL
Std.6	300 µL of Std. 5 (150 pg /ml)	300 µL	75 pg/mL
Std.7	300 µL of Std. 6 (75 pg /ml)	300 µL	37.5 pg/mL
Blank	-	300 µL	0 pg/mL

Note: Do not use a Repeating pipette. Change tips for every dilution. Wet tip with Dilution Buffer before dispensing.

Sample Preparation

Dilute samples with **Dilution Buffer**.

- Conditioned media require neat to appropriate dilution.

Assay Procedure

1. Remove the appropriate number of microtiter wells from the foil pouch and place them into the well holder. Return any unused wells to the foil pouch, refold, seal with tape and store at 4°C.
2. Dilute samples with **Dilution Buffer**. (See "Sample Preparation" above.)
3. Pipette **100 µL** of **Standard Solutions (Std1-Std7, Blank)** and **conditioned media of cultured cells** or **diluted samples** in duplicates, into the appropriate wells.
4. Incubate the plate **at room temperature (ca.25°C) for 1.5 hours, shaking at ca. 300 rpm on an orbital microplate shaker.**
5. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
6. Add **100 µL** of **HRP conjugated Detection Antibody** into each well.
7. Incubate the plate **at room temperature (ca.25°C) for 1.5 hours, shaking at ca. 300 rpm on an orbital microplate shaker.**
8. Wash 4-times by filling each well with Wash Buffer (350 µL) using a squirt bottle, multi-channel pipette, manifold dispenser or microplate washer.
9. Add **100 µL** of **Substrate Reagent**. Avoid exposing the microtiter plate to direct sunlight. Covering the plate with *e.g.* aluminum foil is recommended. Return Substrate Reagent to 4°C immediately after the necessary volume is removed.
10. Incubate the plate **at room temperature (ca.25°C) for 10-20 minutes, shaking at ca. 300 rpm on an orbital microplate shaker.** The incubation time may be extended up to 30 minutes if the reaction temperature is below than 20°C.
11. Add **100 µL** of **Stop Solution** to each well in the same order as the previously added Substrate Reagent.
12. Measure absorbance in each well using a spectrophotometric microplate reader at dual wavelengths of 450/540 nm. Dual wavelengths of 450/550 or 450/595 nm can also be used. Read the microplate at 450 nm if only a single wavelength can be used. Wells must be read within 30 minutes of adding the Stop Solution.

Note-1: Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

Note-2: Reliable standard curves are obtained when either O.D. values do not exceed 0.25 units for the blank (zero concentration), or 3.0 units for the highest standard concentration.

Note-3: If the microplate reader is not capable of reading absorbance greater than the absorbance of the highest standard, perform a second reading at 405 nm. A new standard curve, constructed using the values measured at 405 nm, is used to determine human soluble LOX-1 concentration of off-scale samples. The readings at 405 nm should not replace the on-scale readings at 450 nm.

Calculations

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Plot the optical density for the standards versus the concentration of the standards and draw the best curve. The data can be linearized by using log/log paper and regression analysis may be applied to the log transformation. To determine the human soluble LOX-1 concentration of each sample, first find the absorbance value on the y-axis and extend a horizontal line to the standard curve. At the point of intersection, extend a vertical line to the x-axis and read the corresponding human soluble LOX-1 concentration. If the samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

1. The dose-response curve of this assay fits best to a sigmoidal 4-parameter logistic equation. The results of unknown samples can be calculated with any computer program having a 4-parameter logistic function. It is important to make an appropriate mathematical adjustment to accommodate for the dilution factor.
2. Most microtiter plate readers perform automatic calculations of analyte concentration. The calibration curve is constructed by plotting the absorbance (Y) of calibrators versus log of the known concentration (X) of calibrators, using the four-parameter function. Alternatively, the logit log function can be used to linearize the calibration curve (i.e. logit of absorbance (Y) is plotted versus log of the known concentration (X) of calibrators).

Measurement Range

The measurement range is 37.5 pg/mL to 2,400 pg/mL. Any sample reading higher than the highest standard should be diluted with Dilution Buffer in higher dilution and re-assayed. Dilution factors need to be taken into consideration in calculating the human soluble LOX-1 concentration.

Troubleshooting

1. All samples and standards should be assayed in duplicate, using the protocol described in the **Detailed Protocol**. Incubation times or temperatures significantly different from those specified may give erroneous results.
2. Poor duplicates, accompanied by elevated values for wells containing no sample, indicate insufficient washing. If all instructions in the **Detailed Protocol** were followed accurately, such results indicate a need for washer maintenance.
3. Overall low signal may indicate that desiccation of the plate has occurred between the final wash and addition of Substrate Reagent. Do not allow the plate to dry out. Add Substrate Reagent immediately after wash.

Reagent Stability

All of the reagents included in the MBL Research Product **CircuLex Human soluble LOX-1/OLR1 ELISA Kit** have been tested for stability. Reagents should not be used beyond the stated expiration date. Upon receipt, kit reagents should be stored at 4°C, except the reconstituted Human soluble LOX-1 Standard must be stored below -70°C. Coated assay plates should be stored in the original foil bag sealed by the zip lock and containing a desiccant pack.

Assay Characteristics


1. Sensitivity

The limit of detection (defined as such a concentration of human soluble LOX-1 giving absorbance higher than mean absorbance of blank* plus three standard deviations of the absorbance of blank: A blank + 3SD blank) is better than 12.4 pg/mL of sample.

* Dilution Buffer was pipetted into blank wells.

Thirty two assays were evaluated and the minimum detectable dose (MDD) of human soluble LOX-1. The mean MDD was 12.4 pg/mL. The MDD was determined by adding three standard deviations to the mean optical density value of thirty two zero standard replicates and calculating the corresponding concentration.

Typical Standard Curve



2. Precision

Intra-assay Precision (Precision within an assay)

Three samples of known concentration were tested sixteen times on one plate to assess intra-assay precision.


- Intra-assay (Within-Run, n=3) CV=2.5-2.7 %

Human soluble LOX-1 concentration (ng/mL)

	Sample 1	Sample 2	Sample 3
1	1.25	0.88	0.73
2	1.26	0.87	0.73
3	1.29	0.89	0.74
4	1.31	0.88	0.73
5	1.29	0.89	0.75
6	1.25	0.86	0.73
7	1.26	0.87	0.71
8	1.31	0.90	0.73
9	1.23	0.83	0.69
10	1.24	0.84	0.70
11	1.24	0.84	0.71
12	1.25	0.86	0.74
13	1.30	0.88	0.74
14	1.24	0.86	0.77
15	1.23	0.84	0.72
16	1.33	0.86	0.73
MAX.	1.33	0.90	0.77
MIN.	1.23	0.83	0.69
MEAN	1.27	0.86	0.73
S.D.	0.03	0.02	0.02
C.V.	2.5%	2.5%	2.7%


3. Linearity

To assess the linearity of the assay, samples containing and/or spiked with high concentrations of soluble LOX-1 were serially diluted with the Dilution Buffer to produce samples with values within the dynamic range of the assay.



Example of Test Results

Fig.1 Concentration of soluble LOX-1 in culture supernatant of THP-1 cell (human acute monocytic leukemia cell line) after 50 nM PMA (phorbol 12-myristate 13-acetate) treatment for indicated times.



References

1. Sawamura T, Kume N, Aoyama T, Moriwaki H, Hoshikawa H, Aiba Y, Tanaka T, Miwa S, Katsura Y, Kita T, Masaki T. (1997) *Nature* 386: 73-77
2. Cominacini L, Pasini AF, Garbin U, Davoli A, Tosetti ML, Campagnola M, Rigoni A, Pastorino AM, Lo Cascio V, Sawamura T. (2000) *J Biol Chem.* 275: 12633–12638.
3. Cominacini L, Rigoni A, Pasini AF, Garbin U, Davoli A, Campagnola M, Pastorino AM, Lo Cascio U, Sawamura T. (2001) *J Biol Chem.* 276: 13750–13755.
4. Kume N, Murase T, Moriwaki H, Aoyama T, Sawamura T, Masaki T, Kita T. (1998) *Circ Res.* 83: 322–327.
5. Draude G and Lorenz R. L. (2000) *Am J Physiol Heart Circ Physiol.* 278: H1042–H1048.
6. Nagase M, Hirose S, Sawamura T, Masaki T, Fujita T. (1997) *Biochem Biophys Res Commun.* 237: 496–498.
7. Chen M, Nagase M, Fujita T, Narumiya S, Masaki T, Sawamura T. (2001) *Biochem Biophys Res Commun.* 287: 962–968.
8. Chen M, Kakutani M, Minami M, Kataoka H, Kume N, Narumiya S, Kita T, Sawamura T. (2000) *Arterioscler Thromb Vasc Biol.* 20: 1107–1115.
9. Kataoka H, Kume N, Miyamoto S, Minami M, Moriwaki H, Murase T, Sawamura T, Masaki T, Hashimoto N, Kita T. (1999) *Circulation* 99: 3110–3117.
10. Hayashida K, Kume N, Murase T, Minami M, Nakagawa D, Inada T, Tanaka M, Ueda A, Kominami G, Kambara H, Kimura T, Kita T. (2005) *Circulation* 112: 812-818.

For more information, please visit our web site.

<https://ruo.mbl.co.jp/>

MANUFACTURED BY

MBL A JSR Life Sciences Company
MEDICAL & BIOLOGICAL LABORATORIES CO., LTD.

URL: <https://ruo.mbl.co.jp>

E-mail: support@mbi.co.jp

CycLex/CircuLex products are supplied for research use only. CycLex/CircuLex products and components thereof may not be resold, modified for resale, or used to manufacture commercial products without prior written approval from MBL. To inquire about licensing for such commercial use, please contact us via email.