For Research Use Only.
Not for use in diagnostic procedures.

MBL
A JSR Life Sciences Company

MONOCLONAL ANTIBODY

Anti-HLA-A24 (Human) mAb-FITC

Code No.	Clone	Subclass	Quantity	Concentration
K0208-4	17A10	Mouse IgG2b	100 µL	500 µg/mL

BACKGROUND: HLA (human leukocyte antigen)-A24 is a class I MHC antigen. HLA-A24 is the most frequent HLA class I molecule in Asian populations, present in approximately ~70% of the Japanese population. HLA-A24 is also found in approximately 35% of the Indian population and 19% of Caucasians. HLA antigens may play a role in genetic susceptibility to disease.

SOURCE: This antibody was purified from hybridoma (clone 17A10) supernatant using protein A agarose. This hybridoma was established by fusion of mouse myeloma cell Sp2/0 with C57BL/6 Tg mouse splenocyte immunized with the human recombinant HLA-A24.

FORMULATION: 50 µg IgG in 100 µL volume of PBS containing 1% BSA and 0.1% ProClin 150.

STORAGE: This antibody solution is stable for one year from the date of purchase when stored at 4°C.

REACTIVITY: This antibody reacts with HLA-A24 on Flow cytometry.
Note: It was reported that this clone 17A10 cross-reacted to HLA-B27 and some indeterminate HLA. Although HLA-B27 population is so small in Japanese, about 20% of tested population in our laboratories reacted to this antibody as false-positive. To ensure your experiment, you should confirm HLA genotyping.

APPLICATION:
Flow cytometry; 10 µg/mL (final concentration)
*Please refer to the data sheet (MBL, code no. K0208-3) for other applications.

Detailed procedure is provided in the following **PROTOCOLS**.

SPECIES CROSS REACTIVITY:

Species	Human	Mouse	Rat
Cells	LCL721	Not tested	Not tested
Reactivity on FCM	+		

INTENDED USE:
For Research Use Only. Not for use in diagnostic procedures.

RELATED PRODUCTS:
Other related antibodies and kits are also available.
Please visit our website at https://ruo.mbl.co.jp/.

REFERENCES:
1) Masamoto, I., *et al., Leuk. Lymphoma* **57**, 685-91 (2015). [FCM]
2) Kobayashi, E., *et al., Nat. Med.* **19**, 1542-1546 (2013) [FCM]
3) Kozako, T., *et al., J. Immunol.* **177**, 5718-5726 (2006)
4) Lutz, C. T., *et al., J. Immunol*. **153**, 4099-4110 (1994)
5) Tahara, T., *et al., Immunogenetics* **32**, 351-360 (1990)

Clone 17A10 is used in these references.



Flow cytometric analysis of HLA-A24 expression on LCL721 cells (right) and Jurkat cells (left). *Open histogram indicates the reaction of isotypic control to the cells. Shaded histograms indicate the reaction of K0208-4 to the cells.*

The descriptions of the following protocols are examples. Each user should determine the appropriate condition.

PROTOCOLS:
Flow cytometric analysis for floating cells
We usually use Fisher tubes or equivalents as reaction tubes for all steps described below.
1) Wash the cells 3 times with washing buffer [PBS containing 2% fetal calf serum (FCS) and 0.09% NaN_3].
*Azide may react with copper or lead in plumbing system to form explosive metal azides. Therefore, always flush plenty of water when disposing materials containing azide into drain.
2) Resuspend the cells with washing buffer ($5x10^6$ cells/mL).

3) Add 50 μL of the cell suspension into each tube, and centrifuge at 500 x g for 1 minute at room temperature (20~25°C). Remove supernatant by careful aspiration.
4) Add 20 μL of Clear Back (human Fc receptor blocking reagent, MBL, code no. MTG-001) to the cell pellet after tapping. Mix well and incubate for 10 minutes at room temperature.
5) Add 20 μL of the primary antibody diluted with the washing buffer as suggested in the **APPLICATION**. Mix well and incubate for 15 minutes at room temperature.
6) Add 1 mL of the washing buffer followed by centrifugation at 500 x g for 1 minute at room temperature. Remove supernatant by careful aspiration.
7) Resuspend the cells with 500 μL of the washing buffer and analyze by a flow cytometer.

(Positive control for Flow cytometry; LCL721)

Flow cytometric analysis for whole blood cells

We usually use Fisher tubes or equivalents as reaction tubes for all steps described below.

1) Add 50 μL of the primary antibody at the concentration of as suggested in the **APPLICATION** diluted with the washing buffer [PBS containing 2% fetal calf serum (FCS) and 0.09% NaN_3] into each tube.
2) Add 50 μL of whole blood into each tube. Mix well, and incubate for 30 minutes at room temperature (20~25°C).
3) Add 1 mL of washing buffer followed by centrifugation at 500 x g for 1 minute at room temperature. Remove supernatant by careful aspiration.
4) Lyse with OptiLyse C (for analysis on Beckman Coulter instruments) or OptiLyse B (for analysis on BD instruments), using the procedure recommended in the respective package inserts.
5) Add 1 mL of H_2O to each tube and incubate for 10 minutes at room temperature.
6) Centrifuge at 500 x g for 1 minute at room temperature. Remove supernatant by careful aspiration.
7) Add 1 mL of washing buffer followed by centrifugation at 500 x g for 1 minute at room temperature. Remove supernatant by careful aspiration.
8) Resuspend the cells with 500 μL of the washing buffer and analyze by a flow cytometer.